

-----  
Midmark M11 - v1.0.5  
Total Cycles: 849

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2022 09:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.5 F 0.0  
2:00 165.9 F 2.3  
4:00 183.9 F 6.6  
6:00 206.8 F 12.3  
8:00 222.9 F 8.0  
10:00 235.2 F 11.4  
12:00 241.0 F 11.8  
14:00 240.7 F 11.0  
16:00 254.4 F 18.0  
18:00 266.3 F 25.3  
20:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.8 F 29.8  
2:00 271.5 F 29.4  
3:00 272.8 F 29.9  
4:00 271.6 F 29.2  
  
Min 271.4 F 29.2  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
5:00 232.0 F 0.0  
10:00 215.8 F 0.0  
15:00 194.8 F 0.0  
20:00 185.9 F 0.0  
25:00 185.0 F 0.0  
30:00 183.2 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

03/01/2022 10:09 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02