
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000091
12:45 28/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 4:41
136.0 C 221kPa 4:41
136.4 C 226kPa 5:11
136.4 C 226kPa 5:41
136.3 C 225kPa 6:11
136.1 C 223kPa 6:41
136.2 C 224kPa 7:11
136.4 C 225kPa 7:41
136.3 C 225kPa 8:11
136.3 C 225kPa 8:11
VENTING 8:11
AIR DRYING 9:06
CYCLE COMPLETE 24:06

Digital Signature #
F463EB3141E7944C
