
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001309
7:17 02/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:59
STERILIZING 4:13
136.2 C 221kPa 4:13
136.2 C 222kPa 4:43
136.2 C 222kPa 5:13
136.4 C 223kPa 5:43
136.2 C 220kPa 6:13
136.3 C 221kPa 6:43
136.4 C 222kPa 7:13
136.4 C 222kPa 7:43
136.4 C 222kPa 7:43
VENTING 7:43
AIR DRYING 8:37
CYCLE COMPLETE 8:47

Digital Signature #
B856C43A66E44525
