

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004021  
7:51 15/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:46  
STERILIZING 3:40  
136.0 C 220kPa 3:40  
136.2 C 221kPa 4:10  
136.2 C 223kPa 4:40  
136.3 C 222kPa 5:10  
136.1 C 218kPa 5:40  
136.2 C 218kPa 6:10  
135.9 C 220kPa 6:40  
136.2 C 222kPa 7:10  
136.2 C 222kPa 7:10  
VENTING 7:11  
AIR DRYING 8:00  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
833ED318C66E8E0F  
-----