
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007478
11:50 03/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:16
136.0 C 220kPa 4:16
136.0 C 223kPa 4:46
136.4 C 226kPa 5:16
136.4 C 225kPa 5:46
136.4 C 223kPa 6:16
136.4 C 223kPa 6:46
136.4 C 223kPa 7:16
136.4 C 224kPa 7:46
136.4 C 224kPa 7:46
VENTING 7:46
AIR DRYING 8:44
CYCLE COMPLETE 8:53

Digital Signature #
D2E9DB6DAE708696
