

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2656

Name: RITTER2  
Sterilizer ID: V2313228  
07/17/2023 02:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.1 F 0.0  
2:00 149.6 F 0.7  
4:00 167.4 F 3.6  
6:00 193.2 F 7.4  
8:00 203.3 F 11.6  
10:00 219.9 F 5.8  
12:00 228.6 F 10.4  
14:00 234.1 F 8.7  
16:00 242.5 F 13.0  
18:00 242.2 F 12.0  
20:00 254.5 F 18.2  
22:00 264.9 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.6 F 29.8  
2:00 271.7 F 29.1  
3:00 271.2 F 29.3  
4:00 272.5 F 29.8

Min 271.0 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 216.8 F 0.0  
10:00 219.7 F 0.0  
15:00 197.5 F 0.0  
20:00 197.4 F 0.0  
25:00 208.8 F 0.0  
30:00 216.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:59  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:02:30  
CYCLE COMPLETE

07/17/2023 03:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_