
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007576
13:10 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:54
STERILIZING 3:59
136.4 C 224kPa 3:59
136.3 C 224kPa 4:29
136.3 C 220kPa 4:59
136.4 C 221kPa 5:29
136.4 C 221kPa 5:59
136.1 C 224kPa 6:29
136.2 C 219kPa 6:59
136.2 C 218kPa 7:29
136.2 C 218kPa 7:29
VENTING 7:29
AIR DRYING 8:23
CYCLE COMPLETE 8:50

Digital Signature #
8E26CAD60E5E3547
