
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005094
14:43 09/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:25
PRESSURIZING 1:28
STERILIZING 2:37
136.4 C 221kPa 2:37
136.2 C 219kPa 3:07
136.4 C 220kPa 3:37
136.2 C 218kPa 4:07
136.3 C 222kPa 4:37
136.4 C 222kPa 5:07
136.3 C 220kPa 5:37
136.2 C 222kPa 6:07
136.2 C 222kPa 6:07
VENTING 6:07
AIR DRYING 6:56
CYCLE COMPLETE 8:41

Digital Signature #
82523C493E4B1557
