

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001708  
14:14 08/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:09  
STERILIZING 4:10  
136.0 C 220kPa 4:10  
136.3 C 224kPa 4:40  
136.4 C 225kPa 5:10  
136.1 C 223kPa 5:40  
136.4 C 224kPa 6:10  
136.4 C 225kPa 6:40  
136.3 C 224kPa 7:10  
136.2 C 222kPa 7:40  
136.2 C 222kPa 7:40  
VENTING 7:40  
AIR DRYING 8:36  
CYCLE COMPLETE 8:51

-----  
Digital Signature #  
753D5B2F5C89F7FD  
-----