

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4422

Name: RITTER 1  
Sterilizer ID: V2178198  
04/03/2024 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.2 F 0.0  
2:00 69.6 F 0.3  
4:00 79.2 F 1.2  
6:00 107.9 F 3.2  
8:00 133.8 F 6.0  
10:00 163.0 F 9.3  
12:00 196.9 F 12.7  
14:00 203.5 F 3.6  
16:00 212.3 F 6.8  
18:00 222.5 F 10.3  
20:00 231.2 F 8.9  
22:00 235.6 F 10.1  
24:00 236.4 F 9.0  
26:00 245.5 F 13.2  
28:00 256.8 F 19.2  
30:00 266.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.0  
2:00 271.5 F 28.7  
3:00 272.2 F 29.0  
4:00 271.5 F 28.7

Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 211.6 F 0.0  
10:00 200.8 F 0.0  
15:00 191.9 F 0.0  
20:00 206.3 F 0.0  
25:00 208.7 F 0.0  
30:00 206.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:33  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

04/03/2024 09:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_