
Midmark M11 - v1.0.5
Total Cycles: 3944

Name: RITTER2
Sterilizer ID: V2313228
09/13/2024 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 145.1 F 1.4
4:00 162.0 F 4.3
6:00 187.4 F 8.1
8:00 205.6 F 12.6
10:00 217.0 F 5.9
12:00 228.5 F 10.3
14:00 232.5 F 8.5
16:00 241.2 F 12.3
18:00 240.2 F 10.8
20:00 252.3 F 16.9
22:00 263.1 F 23.3
24:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.7
2:00 271.8 F 28.9
3:00 271.8 F 29.5
4:00 272.4 F 29.5

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 215.5 F 0.0
10:00 201.7 F 0.0
15:00 196.4 F 0.0
20:00 192.1 F 0.0
25:00 194.5 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

09/13/2024 12:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____