
Midmark M11 - v1.0.5
Total Cycles: 1739

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 82.7 F 0.0
4:00 117.8 F 0.0
6:00 162.3 F 0.0
8:00 165.6 F 0.6
10:00 190.2 F 0.7
12:00 209.4 F 3.0
14:00 223.1 F 6.4
16:00 231.9 F 10.1
18:00 237.2 F 9.9
20:00 241.4 F 11.6
22:00 238.4 F 9.8
24:00 237.6 F 9.0
26:00 250.1 F 15.5
28:00 261.7 F 22.2
30:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.5 F 29.3
2:00 271.6 F 28.7
3:00 272.5 F 29.3
4:00 271.7 F 28.6

Min 271.6 F 28.5
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 207.0 F 0.0
10:00 226.2 F 0.0
15:00 225.3 F 0.0
20:00 220.8 F 0.0
25:00 219.3 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

11/17/2022 09:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____