
Midmark M11 - v1.0.0
Total Cycles: 1535

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2022 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 76.3 F 0.0
4:00 87.4 F 0.0
6:00 124.2 F 1.7
8:00 155.7 F 4.5
10:00 175.9 F 8.2
12:00 192.0 F 12.0
14:00 207.9 F 3.7
16:00 217.2 F 7.4
18:00 228.3 F 11.4
20:00 228.9 F 6.8
22:00 241.2 F 12.3
24:00 239.1 F 10.2
26:00 251.8 F 16.5
28:00 262.9 F 22.9
30:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.5 F 28.4
3:00 272.7 F 29.4
4:00 271.9 F 28.7

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 209.1 F 0.0
10:00 214.0 F 0.0
15:00 221.0 F 0.0
20:00 215.5 F 0.0
25:00 213.1 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:18
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

01/19/2022 09:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____