
Midmark M11 - v1.0.5
Total Cycles: 604

Name: RITTER2
Sterilizer ID: V2313228
12/06/2021 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 184.3 F 3.5
4:00 209.1 F 9.5
6:00 224.6 F 7.8
8:00 240.1 F 12.2
10:00 238.4 F 10.7
12:00 237.7 F 10.0
14:00 252.3 F 17.1
16:00 265.0 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.7 F 29.7
2:00 271.2 F 29.1
3:00 272.4 F 29.8
4:00 271.5 F 29.0

Min 271.2 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 232.5 F 0.0
10:00 219.8 F 0.0
15:00 226.0 F 0.0
20:00 224.6 F 0.0
25:00 227.3 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 3:16
HEATING: 17:46
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

12/06/2021 02:31 PM
ACCEPTED: X
REJECTED:
Operator: U02