

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4332

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2024 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.9 F 0.0  
2:00 156.2 F 1.7  
4:00 154.3 F 4.4  
6:00 175.6 F 7.9  
8:00 206.2 F 12.6  
10:00 217.3 F 6.3  
12:00 230.8 F 10.8  
14:00 233.3 F 6.8  
16:00 240.3 F 12.1  
18:00 238.0 F 9.8  
20:00 250.2 F 15.8  
22:00 261.1 F 22.1  
24:00 271.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 271.8 F 29.0  
2:00 271.8 F 29.1  
3:00 271.8 F 29.0  
4:00 271.7 F 28.9

Min 271.6 F 28.9  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 223.8 F 0.0  
10:00 232.1 F 0.0  
15:00 222.2 F 0.0  
20:00 222.7 F 0.0  
25:00 229.4 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 24:32  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:13  
CYCLE COMPLETE

02/26/2024 11:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_