
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007747
13:22 14/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:45
STERILIZING 3:54
136.4 C 224kPa 3:54
136.1 C 223kPa 4:24
136.3 C 224kPa 4:54
136.1 C 219kPa 5:24
136.4 C 224kPa 5:54
136.4 C 224kPa 6:24
136.2 C 224kPa 6:54
136.4 C 221kPa 7:24
136.4 C 221kPa 7:24
VENTING 7:24
AIR DRYING 8:13
CYCLE COMPLETE 38:13

Digital Signature #
A3FC1976EEC37AE8
