
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005362
8:07 18/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:36
STERILIZING 3:44
136.4 C 220kPa 3:44
136.4 C 221kPa 4:14
136.4 C 222kPa 4:44
136.4 C 222kPa 5:14
136.3 C 222kPa 5:44
136.4 C 223kPa 6:14
136.1 C 219kPa 6:44
136.4 C 223kPa 7:14
136.4 C 223kPa 7:14
VENTING 7:14
AIR DRYING 8:01
CYCLE COMPLETE 8:10

Digital Signature #
2B6418F316E7482B
