
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006089
7:32 10/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:12
STERILIZING 4:13
136.5 C 221kPa 4:13
136.2 C 222kPa 4:43
136.0 C 217kPa 5:13
136.4 C 222kPa 5:43
136.4 C 223kPa 6:13
136.4 C 221kPa 6:43
136.4 C 219kPa 7:13
136.4 C 221kPa 7:43
136.4 C 221kPa 7:43
VENTING 7:43
AIR DRYING 8:32
CYCLE COMPLETE 9:01

Digital Signature #
4837C14C1969AAAB
