
Midmark M11 - v1.0.5
Total Cycles: 4316

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.0 F 0.0
2:00 98.5 F 0.0
4:00 108.0 F 0.6
6:00 112.9 F 3.1
8:00 145.3 F 6.2
10:00 166.4 F 9.8
12:00 193.1 F 13.5
14:00 210.0 F 4.7
16:00 219.2 F 8.0
18:00 230.0 F 11.7
20:00 228.1 F 7.4
22:00 239.5 F 12.3
24:00 237.7 F 10.1
26:00 249.1 F 15.7
28:00 259.3 F 21.5
30:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.6 F 30.0
2:00 271.5 F 29.5
3:00 272.4 F 30.1
4:00 271.6 F 29.4

Min 271.5 F 29.3
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 199.4 F 0.0
10:00 204.6 F 0.0
15:00 194.8 F 0.0
20:00 192.3 F 0.0
25:00 202.2 F 0.0
30:00 203.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

01/30/2025 11:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____