
Midmark M11 - v1.0.5
Total Cycles: 4163

Name: RITTER2
Sterilizer ID: V2313228
11/27/2024 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 126.0 F 0.0
4:00 148.7 F 0.3
6:00 184.4 F 2.7
8:00 204.0 F 6.8
10:00 222.7 F 11.5
12:00 227.2 F 7.5
14:00 240.4 F 13.1
16:00 240.1 F 11.3
18:00 239.6 F 10.6
20:00 253.7 F 17.7
22:00 265.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.9
3:00 272.6 F 29.6
4:00 271.7 F 28.9

Min 271.4 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 231.2 F 0.0
10:00 212.0 F 0.0
15:00 186.4 F 0.0
20:00 181.6 F 0.0
25:00 179.7 F 0.0
30:00 178.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

11/27/2024 06:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____