

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1932

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2022 01:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
2:00 182.4 F 3.7  
4:00 192.5 F 8.3  
6:00 211.9 F 13.4  
8:00 226.0 F 8.0  
10:00 238.4 F 13.3  
12:00 238.2 F 10.3  
14:00 235.5 F 8.0  
16:00 249.6 F 14.9  
18:00 261.8 F 21.8  
20:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 28.8  
2:00 271.5 F 28.1  
3:00 272.6 F 28.9  
4:00 271.9 F 28.3  
  
Min 271.3 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 226.1 F 0.0  
10:00 215.1 F 0.0  
15:00 227.8 F 0.0  
20:00 232.3 F 0.0  
25:00 229.0 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

05/11/2022 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_