
Midmark M11 - v1.0.5
Total Cycles: 145

Name: RITTER2
Sterilizer ID: V2313228
06/28/2021 07:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.1 F 0.0
2:00 76.6 F 0.0
4:00 103.5 F 0.1
6:00 92.9 F 0.2
8:00 122.4 F 2.0
10:00 141.6 F 4.7
12:00 166.9 F 7.6
14:00 188.4 F 10.8
16:00 204.1 F 7.2
18:00 212.7 F 7.4
20:00 224.3 F 11.3
22:00 231.5 F 8.0
24:00 240.7 F 13.3
26:00 241.9 F 12.0
28:00 253.0 F 17.7
30:00 262.6 F 23.5
32:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 30.0
1:00 272.0 F 30.1
2:00 271.9 F 30.3
3:00 272.3 F 29.9
4:00 272.4 F 30.8

Min 271.0 F 29.9
Max 272.6 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 228.9 F 0.0
10:00 228.9 F 0.0
15:00 229.1 F 0.0
20:00 229.2 F 0.0
25:00 227.1 F 0.0
30:00 223.8 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 33:04
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:25
CYCLE COMPLETE

06/28/2021 08:43 AM
ACCEPTED: X
REJECTED:
Operator: U01