

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1988

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 01:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
2:00 188.8 F 4.4  
4:00 201.2 F 9.3  
6:00 219.2 F 7.2  
8:00 233.1 F 10.5  
10:00 232.4 F 7.4  
12:00 243.1 F 11.9  
14:00 244.8 F 12.7  
16:00 257.7 F 19.6  
18:00 269.2 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.9 F 29.5  
2:00 271.8 F 28.8  
3:00 272.4 F 29.5  
4:00 272.7 F 29.3  
  
Min 271.5 F 28.7  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 229.3 F 0.0  
10:00 224.9 F 0.0  
15:00 215.2 F 0.0  
20:00 216.1 F 0.0  
25:00 216.3 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

06/01/2022 02:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_