

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1565

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2022 07:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.7 F 0.0  
2:00 80.6 F 0.0  
4:00 103.7 F 0.0  
6:00 151.7 F 0.0  
8:00 160.2 F 0.0  
10:00 179.9 F 0.1  
12:00 202.4 F 2.4  
14:00 216.8 F 5.9  
16:00 227.1 F 9.5  
18:00 236.7 F 13.5  
20:00 237.0 F 10.3  
22:00 236.8 F 8.8  
24:00 242.8 F 11.4  
26:00 246.3 F 13.6  
28:00 258.2 F 20.1  
30:00 268.9 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.3  
2:00 271.5 F 28.6  
3:00 272.8 F 29.4  
4:00 271.8 F 28.7

Min 271.5 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 193.7 F 0.0  
10:00 207.1 F 0.0  
15:00 220.0 F 0.0  
20:00 226.4 F 0.0  
25:00 223.9 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:08  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:04  
CYCLE COMPLETE

10/05/2022 08:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_