
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001391
15:58 19/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:45
STERILIZING 3:53
136.0 C 220kPa 3:53
136.4 C 224kPa 4:23
136.4 C 224kPa 4:53
136.4 C 224kPa 5:23
136.5 C 224kPa 5:53
136.4 C 224kPa 6:23
136.0 C 221kPa 6:53
136.4 C 224kPa 7:23
136.4 C 224kPa 7:23
VENTING 7:23
AIR DRYING 8:15
CYCLE COMPLETE 8:51

Digital Signature #
8DE0DF99D73418BC
