
Midmark M11 - v1.0.0
Total Cycles: 1632

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 77.4 F 0.0
4:00 85.4 F 0.2
6:00 108.3 F 1.3
8:00 137.9 F 2.8
10:00 169.9 F 4.9
12:00 213.1 F 8.6
14:00 221.7 F 12.7
16:00 224.3 F 5.6
18:00 234.4 F 9.8
20:00 238.1 F 9.2
22:00 242.2 F 11.5
24:00 238.9 F 9.5
26:00 252.5 F 16.1
28:00 263.7 F 22.7
30:00 271.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.1 F 28.2
2:00 272.6 F 28.6
3:00 272.8 F 28.9
4:00 272.0 F 28.7

Min 271.6 F 28.1
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.8 F 0.0
5:00 232.9 F 0.0
10:00 219.4 F 0.0
15:00 222.3 F 0.0
20:00 240.0 F 0.0
25:00 194.3 F 0.0
30:00 175.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

02/15/2022 09:24 AM
ACCEPTED: X
REJECTED:
Operator: U02