
Midmark M11 - v1.0.5
Total Cycles: 1455

Name: RITTER2
Sterilizer ID: V2313228
09/06/2022 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 147.8 F 0.0
4:00 165.0 F 1.8
6:00 189.4 F 5.8
8:00 213.5 F 10.8
10:00 222.7 F 6.3
12:00 236.8 F 12.0
14:00 236.7 F 10.0
16:00 237.2 F 8.7
18:00 250.7 F 16.2
20:00 263.0 F 23.3
22:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.8
2:00 272.4 F 29.3
3:00 271.5 F 28.8
4:00 272.5 F 29.8

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 219.6 F 0.0
10:00 227.4 F 0.0
15:00 229.1 F 0.0
20:00 225.7 F 0.0
25:00 226.2 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

09/06/2022 09:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____