
Midmark M11 - v1.0.5
Total Cycles: 4485

Name: RITTER2
Sterilizer ID: V2313228
04/03/2025 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.0 F 0.0
2:00 157.8 F 2.2
4:00 165.9 F 5.6
6:00 192.3 F 10.0
8:00 220.3 F 7.4
10:00 223.9 F 8.6
12:00 235.4 F 13.2
14:00 235.8 F 10.4
16:00 236.7 F 8.6
18:00 247.2 F 14.8
20:00 258.5 F 21.1
22:00 269.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.1 F 30.4
2:00 271.7 F 29.8
3:00 271.4 F 30.2
4:00 272.4 F 30.1

Min 271.0 F 29.5
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 194.8 F 0.0
10:00 197.7 F 0.0
15:00 207.2 F 0.0
20:00 211.5 F 0.0
25:00 209.1 F 0.0
30:00 206.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:01
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

04/03/2025 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____