
Midmark M11 - v1.0.5
Total Cycles: 2340

Name: RITTER2
Sterilizer ID: V2313228
05/04/2023 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.2 F 0.0
2:00 157.4 F 3.3
4:00 176.8 F 7.3
6:00 199.4 F 12.4
8:00 215.7 F 6.3
10:00 231.7 F 11.6
12:00 233.5 F 9.0
14:00 238.4 F 9.9
16:00 247.5 F 14.5
18:00 260.0 F 21.4
20:00 271.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.4 F 29.0
3:00 272.8 F 29.8
4:00 271.8 F 29.1

Min 271.4 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 203.4 F 0.0
10:00 211.3 F 0.0
15:00 211.3 F 0.0
20:00 232.2 F 0.0
25:00 227.1 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

05/04/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____