
Midmark M11 - v1.0.5
Total Cycles: 235

Name: RITTER2
Sterilizer ID: V2313228
07/22/2021 06:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 73.8 F 0.0
4:00 80.2 F 0.7
6:00 116.8 F 2.9
8:00 147.7 F 5.7
10:00 170.5 F 9.3
12:00 197.8 F 13.4
14:00 206.7 F 5.4
16:00 218.7 F 9.1
18:00 230.1 F 13.1
20:00 233.2 F 9.4
22:00 238.8 F 10.5
24:00 244.2 F 12.8
26:00 255.0 F 18.8
28:00 265.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.6
2:00 271.1 F 28.9
3:00 272.3 F 30.1
4:00 273.2 F 29.9

Min 270.9 F 28.8
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 228.7 F 0.0
10:00 190.0 F 0.0
15:00 176.9 F 0.0
20:00 171.3 F 0.0
25:00 168.6 F 0.0
30:00 164.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 29:52
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:08:01
CYCLE COMPLETE

07/22/2021 07:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____