
Midmark M11 - v1.0.5
Total Cycles: 1269

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 126.0 F 1.1
4:00 146.5 F 4.2
6:00 169.2 F 8.2
8:00 190.2 F 12.8
10:00 223.2 F 6.3
12:00 227.8 F 10.9
14:00 228.4 F 7.1
16:00 242.5 F 12.6
18:00 239.5 F 10.8
20:00 252.0 F 16.9
22:00 262.8 F 23.4
24:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.3 F 29.7
2:00 271.8 F 30.1
3:00 271.7 F 29.4
4:00 272.1 F 30.1

Min 271.1 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 226.1 F 0.0
10:00 223.2 F 0.0
15:00 221.1 F 0.0
20:00 218.5 F 0.0
25:00 217.8 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

07/07/2022 03:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____