
Midmark M11 - v1.0.0
Total Cycles: 1990

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2022 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.0 F 0.0
2:00 185.2 F 4.6
4:00 196.4 F 9.5
6:00 222.9 F 7.4
8:00 233.5 F 10.5
10:00 231.4 F 7.0
12:00 244.2 F 12.3
14:00 244.1 F 12.4
16:00 257.0 F 19.2
18:00 268.4 F 26.3
20:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.4
2:00 271.6 F 28.6
3:00 272.6 F 29.5
4:00 272.3 F 29.0

Min 271.4 F 28.6
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 233.2 F 0.0
10:00 227.4 F 0.0
15:00 217.3 F 0.0
20:00 216.8 F 0.0
25:00 218.2 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

06/01/2022 05:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____