
Midmark M11 - v1.0.0
Total Cycles: 3985

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2023 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.3 F 0.0
2:00 111.3 F 1.1
4:00 134.0 F 3.4
6:00 162.0 F 6.6
8:00 183.9 F 10.3
10:00 210.4 F 5.7
12:00 217.5 F 6.4
14:00 228.0 F 10.5
16:00 233.2 F 7.4
18:00 242.5 F 11.8
20:00 238.8 F 9.8
22:00 251.5 F 15.9
24:00 262.0 F 22.4
26:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.7 F 28.8
2:00 272.7 F 29.2
3:00 271.7 F 28.5
4:00 272.4 F 29.2

Min 270.9 F 28.0
Max 273.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 233.1 F 0.0
10:00 219.1 F 0.0
15:00 226.1 F 0.0
20:00 219.3 F 0.0
25:00 196.8 F 0.0
30:00 185.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 26:10
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

11/06/2023 03:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____