
Midmark M11 - v1.0.0
Total Cycles: 1848

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2022 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 182.1 F 4.1
4:00 203.3 F 9.6
6:00 226.6 F 6.2
8:00 235.4 F 10.4
10:00 233.1 F 7.5
12:00 243.3 F 11.7
14:00 246.0 F 13.0
16:00 258.7 F 19.9
18:00 270.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.4
2:00 272.1 F 28.7
3:00 272.0 F 28.2
4:00 272.2 F 28.9

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 228.0 F 0.0
10:00 225.4 F 0.0
15:00 235.1 F 0.0
20:00 225.8 F 0.0
25:00 226.0 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

04/20/2022 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____