

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001552  
13:49 11/03/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:54  
PRESSURIZING 6:39  
STERILIZING 8:09  
136.4 C 222kPa 8:09  
136.2 C 221kPa 8:40  
136.5 C 223kPa 9:10  
136.5 C 223kPa 9:40  
136.1 C 220kPa 10:10  
136.4 C 221kPa 10:40  
136.3 C 221kPa 11:10  
136.4 C 223kPa 11:40  
VENTING 11:40  
AIR DRYING 12:42  
CYCLE COMPLETE 28:16

-----  
Digital Signature #  
62A0823709E4566C  
-----