
Midmark M11 - v1.0.5
Total Cycles: 1168

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 156.4 F 1.5
4:00 156.3 F 5.3
6:00 180.3 F 9.9
8:00 204.9 F 4.9
10:00 220.7 F 7.7
12:00 231.8 F 12.4
14:00 234.0 F 9.4
16:00 238.0 F 10.0
18:00 246.4 F 14.1
20:00 258.4 F 20.6
22:00 269.1 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.9
2:00 271.8 F 29.1
3:00 271.7 F 29.7
4:00 272.4 F 29.8

Min 271.2 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 215.4 F 0.0
10:00 222.0 F 0.0
15:00 221.2 F 0.0
20:00 204.5 F 0.0
25:00 197.5 F 0.0
30:00 198.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

06/07/2022 01:59 PM
ACCEPTED: X
REJECTED:
Operator: U02