
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002210
14:33 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:23
136.5 C 225kPa 4:23
136.4 C 226kPa 4:53
136.3 C 225kPa 5:23
136.2 C 223kPa 5:53
136.3 C 224kPa 6:23
136.2 C 224kPa 6:53
136.4 C 225kPa 7:23
136.4 C 224kPa 7:53
136.4 C 224kPa 7:53
VENTING 7:53
AIR DRYING 8:52
CYCLE COMPLETE 9:01

Digital Signature #
2382078E31A7C405
