
Midmark M11 - v1.0.0
Total Cycles: 3077

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2023 04:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.9 F 0.0
2:00 183.6 F 3.7
4:00 198.3 F 8.0
6:00 221.2 F 13.0
8:00 231.5 F 8.3
10:00 241.9 F 12.4
12:00 241.8 F 11.4
14:00 239.7 F 10.0
16:00 253.5 F 16.8
18:00 265.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.1 F 28.4
2:00 271.8 F 28.6
3:00 271.9 F 28.2
4:00 271.9 F 28.7

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 237.9 F 0.0
10:00 229.0 F 0.0
15:00 228.4 F 0.0
20:00 225.7 F 0.0
25:00 228.0 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

04/04/2023 05:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____