

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001582  
13:37 09/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:43  
STERILIZING 3:40  
136.0 C 220kPa 3:40  
136.4 C 226kPa 4:10  
136.4 C 226kPa 4:40  
136.1 C 224kPa 5:10  
136.3 C 225kPa 5:40  
136.2 C 223kPa 6:10  
136.4 C 225kPa 6:40  
136.2 C 222kPa 7:10  
136.2 C 222kPa 7:10  
VENTING 7:10  
AIR DRYING 8:00  
CYCLE COMPLETE 8:05

-----  
Digital Signature #  
2FBCE2EBBCD84F43  
-----