

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3793

Name: RITTER2  
Sterilizer ID: V2313228  
07/17/2024 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
2:00 182.9 F 1.7  
4:00 181.9 F 5.2  
6:00 191.9 F 9.6  
8:00 213.4 F 9.5  
10:00 224.6 F 8.3  
12:00 235.6 F 13.1  
14:00 236.8 F 10.5  
16:00 236.3 F 8.6  
18:00 248.2 F 15.1  
20:00 259.2 F 21.2  
22:00 269.2 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.7 F 30.0  
2:00 271.5 F 29.2  
3:00 272.7 F 30.1  
4:00 271.9 F 29.5

Min 271.2 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 227.0 F 0.0  
10:00 222.4 F 0.0  
15:00 220.4 F 0.0  
20:00 218.5 F 0.0  
25:00 232.3 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:02:19  
CYCLE COMPLETE

07/17/2024 11:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_