

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2752

Name: RITTER 1  
Sterilizer ID: V2178198  
01/09/2023 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.2 F 0.0  
2:00 138.4 F 0.0  
4:00 168.1 F 0.6  
6:00 190.0 F 4.7  
8:00 207.8 F 10.0  
10:00 221.7 F 6.8  
12:00 231.8 F 9.6  
14:00 236.3 F 9.6  
16:00 242.9 F 12.3  
18:00 240.8 F 10.9  
20:00 254.1 F 17.5  
22:00 265.4 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.5 F 28.3  
3:00 272.4 F 29.2  
4:00 272.2 F 28.8

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 217.2 F 0.0  
10:00 214.1 F 0.0  
15:00 226.2 F 0.0  
20:00 229.0 F 0.0  
25:00 233.8 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:48  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:51  
CYCLE COMPLETE

01/09/2023 03:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_