
Midmark M11 - v1.0.5
Total Cycles: 2927

Name: RITTER2
Sterilizer ID: V2313228
09/26/2023 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 160.2 F 0.1
4:00 166.5 F 1.2
6:00 187.6 F 5.4
8:00 208.6 F 10.2
10:00 219.2 F 6.6
12:00 231.8 F 10.6
14:00 232.7 F 7.8
16:00 243.9 F 13.2
18:00 243.4 F 12.4
20:00 255.4 F 18.7
22:00 266.4 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.6
2:00 271.6 F 28.8
3:00 272.5 F 29.7
4:00 272.0 F 29.0

Min 271.4 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 210.7 F 0.0
10:00 206.8 F 0.0
15:00 206.8 F 0.0
20:00 203.7 F 0.0
25:00 206.8 F 0.0
30:00 213.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:00
CYCLE COMPLETE

09/26/2023 10:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____