
Midmark M11 - v1.0.0
Total Cycles: 3935

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2023 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 159.7 F 2.5
4:00 192.4 F 5.9
6:00 203.8 F 10.7
8:00 215.1 F 4.5
10:00 227.2 F 9.4
12:00 235.7 F 9.8
14:00 239.4 F 11.3
16:00 237.7 F 9.4
18:00 250.2 F 15.7
20:00 261.6 F 22.0
22:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 271.8 F 28.9
2:00 272.1 F 29.0
3:00 272.4 F 29.2
4:00 272.7 F 29.4

Min 271.6 F 28.9
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 241.9 F 0.0
10:00 234.1 F 0.0
15:00 231.7 F 0.0
20:00 229.8 F 0.0
25:00 226.7 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

10/23/2023 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____