
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001739
11:11 17/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:47
STERILIZING 3:45
136.0 C 220kPa 3:45
136.4 C 225kPa 4:15
136.1 C 223kPa 4:45
136.4 C 222kPa 5:15
136.4 C 223kPa 5:45
136.1 C 222kPa 6:15
136.4 C 225kPa 6:45
136.4 C 224kPa 7:15
136.4 C 224kPa 7:15
VENTING 7:15
AIR DRYING 8:03
CYCLE COMPLETE 9:20

Digital Signature #
880455D88441DC9F
