
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008073
15:50 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:24
STERILIZING 3:19
136.0 C 220kPa 3:19
136.1 C 222kPa 3:49
135.9 C 220kPa 4:19
136.4 C 221kPa 4:49
136.4 C 224kPa 5:19
136.4 C 224kPa 5:49
136.4 C 223kPa 6:19
136.4 C 222kPa 6:49
136.4 C 222kPa 6:49
VENTING 6:49
AIR DRYING 7:43
CYCLE COMPLETE 37:43

Digital Signature #
124A3237EA76F133
