

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1847

Name: RITTER 1  
Sterilizer ID: V2178198  
04/20/2022 10:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
2:00 191.3 F 3.8  
4:00 222.3 F 9.4  
6:00 225.9 F 6.6  
8:00 234.4 F 9.8  
10:00 237.3 F 8.6  
12:00 245.1 F 12.8  
14:00 243.2 F 11.5  
16:00 256.2 F 18.3  
18:00 267.8 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.3 F 28.3  
2:00 271.7 F 28.2  
3:00 272.3 F 28.3  
4:00 271.5 F 28.1  
  
Min 271.4 F 27.8  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 229.7 F 0.0  
10:00 217.1 F 0.0  
15:00 216.3 F 0.0  
20:00 216.4 F 0.0  
25:00 230.3 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 19:18  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:48  
CYCLE COMPLETE

04/20/2022 11:27 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_