
Midmark M11 - v1.0.5
Total Cycles: 980

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 75.7 F 0.0
4:00 88.0 F 0.0
6:00 98.1 F 0.0
8:00 157.3 F 2.2
10:00 176.7 F 5.3
12:00 194.0 F 9.2
14:00 216.8 F 13.0
16:00 224.0 F 6.4
18:00 232.9 F 10.2
20:00 238.3 F 9.2
22:00 240.0 F 11.9
24:00 238.8 F 10.2
26:00 251.4 F 16.5
28:00 262.2 F 22.8
30:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.5
2:00 271.6 F 29.4
3:00 272.4 F 29.3
4:00 271.8 F 29.6

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 197.5 F 0.0
10:00 195.3 F 0.0
15:00 195.8 F 0.0
20:00 203.9 F 0.0
25:00 209.8 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:27
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

04/13/2022 08:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____