
Midmark M11 - v1.0.0
Total Cycles: 3434

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.5 F 0.0
2:00 196.1 F 2.9
4:00 197.3 F 6.9
6:00 215.6 F 12.7
8:00 228.1 F 7.6
10:00 241.1 F 13.1
12:00 239.7 F 10.8
14:00 237.6 F 9.2
16:00 252.0 F 16.2
18:00 264.0 F 23.1
20:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.1
3:00 272.7 F 28.9
4:00 271.9 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 233.1 F 0.0
10:00 227.2 F 0.0
15:00 228.5 F 0.0
20:00 227.3 F 0.0
25:00 229.1 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

06/23/2023 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____