
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000810
14:53 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 5:26
STERILIZING 6:48
136.0 C 220kPa 6:48
136.1 C 221kPa 7:18
136.4 C 223kPa 7:48
136.4 C 223kPa 8:18
136.3 C 222kPa 8:48
136.1 C 221kPa 9:18
136.4 C 222kPa 9:48
136.3 C 219kPa 10:18
136.3 C 219kPa 10:18
VENTING 10:19
AIR DRYING 11:20
CYCLE COMPLETE 36:11

Digital Signature #
2468898C5C7970B3
