

-----  
Midmark M11 - v1.0.5  
Total Cycles: 749

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2022 07:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 83.1 F 0.0  
4:00 110.6 F 0.0  
6:00 112.9 F 0.1  
8:00 152.5 F 0.3  
10:00 178.8 F 3.1  
12:00 195.6 F 6.5  
14:00 208.6 F 10.6  
16:00 218.2 F 6.8  
18:00 229.3 F 8.6  
20:00 239.4 F 13.2  
22:00 239.6 F 10.8  
24:00 237.6 F 9.4  
26:00 251.4 F 16.1  
28:00 263.0 F 22.8  
30:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.2  
2:00 271.5 F 28.5  
3:00 272.8 F 29.4  
4:00 271.9 F 28.5

Min 271.5 F 28.4  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 166.4 F 0.0  
5:00 213.4 F 0.0  
10:00 201.3 F 0.0  
15:00 200.8 F 0.0  
20:00 199.6 F 0.0  
25:00 199.2 F 0.0  
30:00 197.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:15  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:08:07  
CYCLE COMPLETE

02/02/2022 08:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_